


File D:\DATA\912
 Acq. Method : D:\METHODS\1-POS-3MIN.M
 Analysis Method : D:\METHODS\1-POS-3MIN.M
 Sample Info : *** No target masses specified ***
 Method Info : Mobile Phase: A: water(0.01%TFA) B: ACN(0.01%TFA)
 Gradient: 5%-95% B in 1.2min
 Flow Rate: 2.2 ml/min
 Column: Poroshell 120 EC-C18,4.6*30mm,2.7um
 Oven Temperature: 50 C
 Mass Range: 110-800



Integration Results for ADC1 A, ELSD

RetTim	Width	Area	Height	Area%
0.18	0.02	3.92	3.77	4.73
1.44	0.02	75.47	70.75	91.22
2.33	0.02	3.35	2.36	4.05

Integration Results for DAD1 A, Sig=214,4 Ref=off

RetTim	Width	Area	Height	Area%
1.44	0.01	360.42	537.31	100.00


Integration Results for DAD1 B, Sig=254,4 Ref=off

RetTim	Width	Area	Height	Area%
1.44	0.01	386.47	614.38	100.00

LC/MS Report

<<<< POSITIVE SPECTRA >>>>

Ret. Time: 1.45



Ret. Time: 2.35

