


# HPLC Analysis for ARN-509 (Apalutamide)



=====  
Acq. Operator : 1  
Acq. Instrument : Instrument 1 Location : Vial 0  
Injection Date : 4/7/2017 3:01:14 PM  
Acq. Method : D:\CHEM32\1\METHODS\OTHER\A\ACN\131209 A-FULL\_ACID\_10 MIN.M  
Last changed : 12/9/2013 8:15:27 PM by 1  
Analysis Method : D:\CHEM32\1\METHODS\ATHENA 50 MM\SPECIAL M\131223 BICOLL\_ATH-M03\_A.M  
Last changed : 4/7/2016 3:12:46 PM by 1  
(modified after loading)  
Sample Info : A = ACN, B = 0.1% H3PO4/H2O] 10-95% A/B 3.5 min, 95% A/  
B 4.5 min, 95-10% A/B 1 min, 10% A/B 1 min; Athena C18  
(3 µm, 2.1 mm × 100 mm); Flow: 0.4 ml/min, T = 25 °C  
=====



=====  
Area Percent Report  
=====

Sorted By : Signal  
Multiplier: : 1.0000  
Dilution: : 1.0000  
Use Multiplier & Dilution Factor with ISTDs

Signal 1: VWD1 A, Wavelength=254 nm

Peak #	RetTime [min]	Type	Width [min]	Area mAU *s	Height [mAU]	Area %
1	3.925	BB	0.0453	20.86998	7.00540	0.2024
2	4.752	BB	0.0570	1.01470e4	2815.35913	98.4122
3	5.028	BB	0.0457	142.84641	48.42899	1.3854

Totals : 1.03107e4 2870.79352

=====  
\*\*\* End of Report \*\*\*